

EC-Declaration of Conformity

EC – Declaration of Conformity

NovoFine®

30G 8 mm

31G 6 mm

32G 6 mm

32G 4 mm

Uma Prathyusha Gupta N
RA Device Established Products

EC - Declaration of Conformity

We, **Legal Manufacturer/ Market Authorization Holder**

Novo Nordisk A/S
Novo Allé
2880 Bagsværd
Denmark

Being the manufacturer/distributor within the European Economic Area, declare that this Declaration of Conformity is issued under our sole responsibility and covers the following product(s):

Product Name	Item number (4 or 5-number)	Classification	GMDN Code
NovoFine® 30G 8 mm	5-4325-21 5-4327-16	Class IIa	44127
NovoFine® 31G 6 mm	5-4356-16 5-4356-17	Class IIa	44127
NovoFine® 32G 6 mm	5-4314-18 5-4314-17	Class IIa	44127
NovoFine® 32G 4 mm	5-4401-11	Class IIa	44127

manufactured in the below mentioned production facilities:

Nipro Medical Industries Ltd., 2-19-64, Matsubara, Tatebayashi-shi, Gunma, 374-8518, Japan
Nipro (Thailand) Corporation Ltd., 10/2 Moo 8, Bangnomko, Sena, Phra Nakhon Si Ayutthaya, 13110, Thailand
Needle Manufacturing & Sourcing, Stenager Allé, 9800 Hjørring, Denmark

declare that the above is in conformity with the provisions of the Council Directive

European Council Medical Device Directive 93/42/EEC, of 14 June 1993, inclusive amendment 2007/47/EC of 5 September 2007.

The above devices are CE-marked and classified as IIa according to Annex IX, Classification Criteria, rule 6.

The devices have been subject to the conformity procedure laid down in Annex II under the supervision of TÜV SÜD Product Service GmbH, a Notified Body authorized by the German Competent Authority, and carrying the Notified Body number 0123.

Notified Body Address: TÜV SÜD PRODUCT SERVICE GMBH
 PS-NAM1-MUC
 Ridlerstr. 65
 80339 Munchen
 Germany

The following standards have been observed:

EN ISO 13485:2016/AC:2018	Medical devices - Quality management systems – Requirements for regulatory purposes
EN ISO 14971:2019	Medical devices - Application of risk management to medical devices
EN ISO 11608-2:2012	Needle-based injection systems for medical use – Requirements and test methods – Part 2: Needles
EN ISO 14155:2011	Clinical investigation of medical devices for human subjects – good clinical practice
EN 62366-1:2015	Medical devices – Part 1: Application of usability engineering to medical devices
ISO 11607-1: 2019	Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging
ISO 11607-2:2019	Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes
EN ISO 11135:2014/A1:2019	Sterilization of health-care products - Ethylene oxide - Requirements for the development, validation and routine control of a sterilization process for medical devices – Amendment 1: Revision of Annex E, Single batch release (ISO 11135:2014/Amd 1:2018)
EN ISO 11737-1:2006 + AC:2009 (NMS)	Sterilization of medical devices – Microbiological Methods-Part 1: Determination of a population of microorganisms on products
EN ISO 11737-1:2018 (NMI)	Sterilization of medical devices – Microbiological Methods-Part 1: Determination of a population of microorganisms on products

EN ISO 11737-2: 2009 (NTC)	Sterilization of medical devices - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process
ISO 11737-2:2019 (NMI)	Sterilization of health care products - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process
ISO 10993-1:2018	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process
EN ISO 14644-1:2015	Cleanrooms and associated controlled environments – Part 1: Classification of air cleanliness by particle concentration
EN ISO 14644-2:2015	Cleanrooms and associated controlled environments – Part 2: Monitoring to provide evidence of cleanroom performance related to air cleanliness by particle concentration
EN 1041:2008+A1:2013	Information supplied by the manufacturer of medical devices
EN ISO 15223-1:2016	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements

Location: Hjørring

On: 2020.11.25

By:

Rita M Jensen

Name & function of person signing

Director of Quality.

Certificate of Analysis



isk®

Novo Nordisk A/S
Novo Allé
2880 Bagsvaerd
Denmark
Tel +45 4444 8888
Fax +45 4449 0665

NOVOFINE 30G X 8MM
100PCS.

Product No. 7200165
Batch No. 20D06M

Order Number : 7735583-000

Date of Manufacture : 04/2020

Date of Expiry : 03/2025

Supplier: Nipro Thailand Corporation LTD, Thailand

Method of sterility: Gas (ethylene oxide)

Sterility: Complies with the USP ed. 42 <71> and
EN 556.

We hereby declare that the above-mentioned batch
was manufactured and controlled in compliance with
the GMP regulations and approved according to the
quality specifications of Novo Nordisk A/S.

20200923, Hjoerring

LENE JØRGENSEN

Quality Department

The Certificate of Analysis is generated by a computer quality
system udes-an-telec tured _

304888 LNJ

064 HUNGARY

Diszpó: NN8000035362 Vevő: Pharmexpert Europe File név: NN00002015 Egyedi sorszám: NN0000045175 13 másolat
Szállító: UPS Healthcare Hungary Zrt. Oldal/Lap 1 / 1 Anyaglap: NN0274/20
Az eredetiről készült hiteles másolat.
Kizárólag a kék tintával szignált másolat tekinthető hitelesnek! Nyomtatás dátuma: 2021.02.04

CertEx v3.10

